

共用基準範囲を採用した項目一覧

~2016/5/8				2016/5/9~			
検査項目	基準範囲		単位	検査項目	基準範囲		単位
	下限	上限			下限	上限	
WBC	M:3.9	M:9.8	$\times 10^3/\mu$	WBC	3.3	8.6	$\times 10^3/\mu$
	F:3.5	F:9.1					
RBC	M:4.27	M:5.70	$\times 10^6/\mu$	RBC	M:4.35	M:5.55	$\times 10^6/\mu$
	F:3.76	F:5.00			F:3.86	F:4.92	
Hb	M:13.5	M:17.6	g/dl	Hb	M:13.7	M:16.8	g/dl
	F:11.3	F:15.2			F:11.6	F:14.8	
Hct	M:39.8	M:51.8	%	Hct	M:40.7	M:50.1	%
	F:33.4	F:44.9			F:35.1	F:44.4	
MCV	M:82.7	M:101.6	fl	MCV	83.6	98.2	fl
	F:79.0	F:100.0					
MCH	M:28.0	M:34.6	pg	MCH	27.5	33.2	pg
	F:26.3	F:34.3					
MCHC	M:31.6	M:36.6	%	MCHC	31.7	35.3	%
	F:30.7	F:36.6					
PLT	M:13.1	M:36.2	$\times 10^4/\mu$	PLT	158	348	$\times 10^3/\mu$
	F:13.0	F:36.9					
TP	6.7	8.3	g/dl	TP	6.6	8.1	g/dl
ALB	4.0	5.0	g/dl	ALB	4.1	5.1	g/dl
GLB	2.2	3.4	g/dl	GLB	2.2	3.4	g/dl
A/G	1.32	2.23		A/G	1.32	2.23	
UN	8	22	mg/dl	UN	8	20	mg/dl
CRE	0.4	1.10	mg/dl	CRE	M:0.65	M:1.07	mg/dl
					F:0.46	F:0.79	
UA	2.3	8.0	mg/dl	UA	M:3.7	M:7.8	mg/dl
					F:2.6	F:5.5	
Na	138	146	mmol/l	Na	138	145	mmol/L
K	3.6	4.9	mmol/l	K	3.6	4.8	mmol/L
Cl	99	109	mmol/l	Cl	101	108	mmol/L
Ca	9.0	10.6	mg/dl	Ca	8.8	10.1	mg/dl
IP	2.5	4.7	mg/dl	IP	2.7	4.6	mg/dl

~2016/5/8

2016/5/9~

検査項目	基準範囲		単位	検査項目	基準範囲		単位
	下限	上限			下限	上限	
GLU	70	110	mg/dl	GLU	73	109	mg/dl
TG	38	207	mg/dl	TG	M:40 F:30	M:234 F:117	mg/dl
TC	128	256	mg/dl	TC	142	248	mg/dl
HDL-C	41	96	mg/dl	HDL-C	M:38 F:48	M:90 F:103	mg/dl
LDL-C	60	153	mg/dl	LDL-C	65	163	mg/dl
TB	0.3	1.5	mg/dl	TB	0.4	1.5	mg/dl
AST	13	33	U/L	AST	13	30	U/L
ALT	8	42	U/L	ALT	M:10 F:7	M:42 F:23	U/L
LD	119	229	U/L	LD	124	222	U/L
ALP	115	359	U/L	ALP	106	322	U/L
GGT	10	47	U/L	GGT	M:13 F:9	M:64 F:32	U/L
ChE	214	466	U/L	CHE	M:240 F:201	M:486 F:421	U/L
AMY	40	130	U/L	AMY	44	132	U/L
CK	M:62 F:45	M:287 F:163	U/L	CK	M:59 F:41	M:248 F:153	U/L
CRP	0	0.17	mg/dl	CRP	0	0.14	mg/dl
Fe	M:55 F:51	M:163 F:139	μg/dl	Fe	40	188	μg/dl
IgG	870	1700	mg/dl	IgG	861	1747	mg/dl
IgA	110	410	mg/dl	IgA	93	393	mg/dl
IgM	35	220	mg/dl	IgM	M:33 F:50	M:183 F:269	mg/dl
C3	65	135	mg/dl	C3	73	138	mg/dl
C4	13	35	mg/dl	C4	11	31	mg/dl
HbA1c	4.3	5.8	%	HbA1c	4.9	6.0	%

PLTは単位が変更になっています。ご注意ください。

基準範囲を見直した項目一覧

～2016/5/8				2016/5/9～			
検査項目	基準範囲		単位	検査項目	基準範囲		単位
	下限	上限			下限	上限	
P-AMY	7	50	U/L	P-AMY	16	52	U/L
Mg	1.8	2.5	mg/dl	Mg	1.8	2.4	mg/dl
CK-MB	8	17	U/L	CK-MB	0	15	U/L
UIBC	M:126	M:294	μg/dl	UIBC	M:111	M:255	μg/dl
	F:150	F:302			F:137	F:325	
NAG	1.0	7	IU/L	NAG	0	11.5	IU/L
尿中微量アルブミン	1.0	31	mg/L	尿中微量アルブミン	0	30	mg/L
CH50	30	45	CH50/ml	CH50	30	46	CH50/ml
蛋白分画				蛋白分画			
Alb	59.6	71.4	%	Alb	61.7	71.7	%
α-G	1.9	3.4	%	α-G	1.7	2.6	%
α-G	5.7	9.8	%	α-G	5.7	8.9	%
β-G	6.6	10.7	%	β-G	6.6	10.2	%
γ-G	10.3	20.3	%	γ-G	10.8	20.1	%
血液ガス				血液ガス			
pH	7.35	7.45	変更なし	pH	7.35	7.45	
pCO2	35.0	45.0	mmHg	pCO2	M:35	M:48	mmHg
					F:32	F:45	
pO2	75	100	mmHg	pO2	83	108	mmHg
HCO3				HCO3	M:22.2	M:28.3	mmol/L
					F:21.2	F:27.0	
SO2	92.0	98.5	%	SO2	95.0	99.0	%
BE	-2	+2	mmol/L	BE	-7	+2	mmol/L