

Month	Year
5	2023

Day of Month	Non Business Day	Refrigerator 1 Logger ④			Refrigerator 2 Logger ⑥			Refrigerator 4 Logger ⑩			Refrigerator 5 Logger ⑪			Room Logger ⑫			MONITORING (✓)	INITIALS
		Time of reading 24-hr clock	Temperature (°C)	Alert (✓)	NoAlert (✗)	Temperature (°C)	Alert (✓)	NoAlert (✗)	Temperature (°C)	Alert (✓)	NoAlert (✗)	Temperature (°C)	Alert (✓)	NoAlert (✗)	Temperature (°C)	Alert (✓)	NoAlert (✗)	
1		14:25	4.6	✓	✗	4.8	✓	✗	5.3	✓	✗	4.8	✓	✗	22.2	✓	✗	K.S
2		13:40	4.7	✓	✗	5.0	✓	✗	5.2	✓	✗	4.8	✓	✗	21.9	✓	✗	K.S
3	✓	—:																
4	✓	—:																
5	✓	—:																
6	✓	—:																
7	✓	—:																
8		15:00	4.9	✓	✗	5.0	✓	✗	5.2	✓	✗	4.8	✓	✗	22.2	✓	✗	K.S
9		15:10	4.9	✓	✗	5.0	✓	✗	5.3	✓	✗	4.2	✓	✗	22.2	✓	✗	K.S
10		15:05	4.9	✓	✗	5.0	✓	✗	5.2	✓	✗	4.5	✓	✗	22.5	✓	✗	K.S
11		14:05	4.6	✓	✗	4.7	✓	✗	5.1	✓	✗	4.1	✓	✗	22.5	✓	✗	K.S
12		15:00	4.8	✓	✗	5.0	✓	✗	5.2	✓	✗	4.2	✓	✗	22.6	✓	✗	K.S
13	✓	—:																
14	✓	—:																
15		14:25	4.6	✓	✗	4.9	✓	✗	5.1	✓	✗	4.5	✓	✗	22.6	✓	✗	K.S
16		14:35	4.8	✓	✗	4.7	✓	✗	5.0	✓	✗	4.1	✓	✗	22.9	✓	✗	K.S
17		17:15	4.6	✓	✗	4.9	✓	✗	5.1	✓	✗	4.3	✓	✗	22.4	✓	✗	K.S
18		15:00	4.9	✓	✗	4.7	✓	✗	5.0	✓	✗	4.3	✓	✗	22.7	✓	✗	K.S
19		15:05	4.9	✓	✗	4.9	✓	✗	5.0	✓	✗	4.3	✓	✗	22.9	✓	✗	K.S
20	✓	—:																
21	✓	—:																
22		14:30	4.8	✓	✗	5.0	✓	✗	5.0	✓	✗	4.0	✓	✗	22.9	✓	✗	K.S
23		14:50	4.8	✓	✗	5.1	✓	✗	5.1	✓	✗	4.3	✓	✗	22.9	✓	✗	K.S
24		14:50	4.8	✓	✗	4.8	✓	✗	5.2	✓	✗	4.3	✓	✗	22.8	✓	✗	K.S
25		15:20	4.7	✓	✗	4.8	✓	✗	4.9	✓	✗	4.4	✓	✗	22.7	✓	✗	K.S
26		14:25	4.9	✓	✗	5.1	✓	✗	5.1	✓	✗	4.3	✓	✗	22.9	✓	✗	K.S
27	✓	—:																
28	✓	—:																
29		14:40	4.9	✓	✗	5.1	✓	✗	4.9	✓	✗	4.4	✓	✗	23.2	✓	✗	K.S
30		15:50	5.0	✓	✗	4.8	✓	✗	5.1	✓	✗	4.2	✓	✗	22.9	✓	✗	K.S
31		14:45	5.0	✓	✗	4.9	✓	✗	5.0	✓	✗	4.3	✓	✗	22.7	✓	✗	K.S

Refrigerator No.		1	2	4	5	INITIALS
Day of Month	Time of reading 24-hr clock	No abnormality (✗)	No abnormality (✗)	No abnormality (✗)	No abnormality (✗)	
2	13:40	✓	✓	✓	✓	K.S