

Month	Year
8	2023

Day of Month	Non Business Day	Time of reading 24-hr clock	Refrigerator 1 Logger ④			Refrigerator 2 Logger ⑥			Refrigerator 4 Logger ⑩			Refrigerator 5 Logger ⑧			Refrigerator ① Logger ⑮			Room Logger ②			MONITORING (✓)	INITIALS
			Temperature (°C)	Alarm (✓)	No Alarm (✓)	Temperature (°C)	Alarm (✓)	No Alarm (✓)	Temperature (°C)	Alarm (✓)	No Alarm (✓)	Temperature (°C)	Alarm (✓)	No Alarm (✓)	Temperature (°C)	Alarm (✓)	No Alarm (✓)	Temperature (°C)	Alarm (✓)	No Alarm (✓)		
1		15:20	5.2		✓	5.3		✓	5.4		✓	4.4		✓				23.0		✓	✓	K.S
2		14:30	5.2		✓	5.3		✓	5.5		✓	4.5		✓				23.2		✓	✓	K.S
3		15:00	5.0		✓	5.3		✓	5.5		✓	4.3		✓				23.3		✓	✓	K.S
4		15:05	5.2		✓	5.1		✓	5.5		✓	4.5		✓				23.3		✓	✓	K.S
5	✓	___:___																				
6	✓	___:___																				
7		14:45	4.9		✓	5.2		✓	5.5		✓	4.0		✓				23.3		✓	✓	K.S
8		16:45				5.1		✓				4.4		✓	4.5		✓	23.5		✓	✓	K.S
9		11:40				5.3		✓				4.4		✓	5.3		✓	23.5		✓	✓	K.S
10		15:40				5.1		✓				4.5		✓	4.0		✓	23.3		✓	✓	K.S
11	✓	___:___																				
12	✓	___:___																				
13	✓	___:___																				
14		14:20				5.2		✓				4.1		✓	5.3		✓	23.2		✓	✓	K.S
15		14:55				5.2		✓				4.3		✓	4.1		✓	23.4		✓	✓	K.S
16		14:10				5.0		✓				4.3		✓	5.1		✓	23.4		✓	✓	K.S
17		15:35				5.1		✓				4.5		✓	3.6		✓	23.5		✓	✓	K.S
18		15:40				5.0		✓				4.3		✓	3.8		✓	23.4		✓	✓	K.S
19	✓	___:___																				
20	✓	___:___																				
21		15:20				5.2		✓				4.4		✓	3.6		✓	23.5		✓	✓	K.S
22		15:25				5.2		✓				4.6		✓	3.6		✓	23.2		✓	✓	K.S
23		15:10				5.1		✓				4.5		✓	4.3		✓	23.3		✓	✓	K.S
24		15:00				5.2		✓				4.4		✓	4.9		✓	23.4		✓	✓	K.S
25		12:00				5.0		✓				4.3		✓	3.5		✓	23.1		✓	✓	K.S
26	✓	___:___																				
27	✓	___:___																				
28		14:20	5.0		✓	5.1		✓	5.6		✓	4.4		✓	5.1		✓	23.3		✓	✓	K.S
29		15:20	4.9		✓	5.1		✓	5.4		✓	4.6		✓	5.2		✓	23.3		✓	✓	K.S
30		15:50	5.0		✓	5.0		✓	5.4		✓	4.2		✓	4.8		✓	23.4		✓	✓	K.S
31		14:55	4.9		✓	5.0		✓	5.3		✓	4.3		✓	5.7		✓	23.3		✓	✓	K.S

Refrigerator No.		1	2	4	5	6	INITIALS
Day of Month	Time of reading 24-hr clock	No abnormality (✓)	No abnormality (✓)	No abnormality (✓)	No abnormality (✓)	No abnormality (✓)	
31	15:00	✓	✓	✓	✓	✓	K.S

8/8 11:10 ④⑩

電池残量少なため 冷蔵庫から出した。

④のかわり ⑮

⑩のかわり ⑬ ⑫元の冷蔵庫5へ

⑮⑬ 11:30 測定キコフ開始

8/8 ④⑩ 修理から戻ってきたため記録開始。(8:00-)

④を冷蔵庫①

⑩を冷蔵庫④へ戻した。